EMC EMISSION - TEST REPORT

Report Number	:	**64.711.17.02847.01– (E)**	Date of Issue:	2017-07-25

Model / Serial No.	:	DW-40L420F, DW-30L420F
Product Type	:	Deep Freezer
Applicant	:	Haier Medical and Laboratory Products Co., Ltd.
Manufacturer	:	Haier Medical and Laboratory Products Co., Ltd.
License holder	:	Haier Medical and Laboratory Products Co., Ltd.
Address	:	Haier Industrial Park, Economic Technology Development Zone, Qingdao, 266510, P.R. China
Test Result	:	■ **Positive** ☐ **Negative**
Total pages including Appendices	:	**30**

TÜV SÜD Certification and Testing (China) Co., Ltd. Guangzhou Branch is a subcontractor to TÜV SÜD Product Service, GmbH according to the principles outlined in ISO/IEC Guide 25 and EN 45001.

TÜV SÜD Certification and Testing (China) Co., Ltd. Guangzhou Branch reports apply only to the specific samples tested under stated test conditions. Construction of the actual test samples has been documented. It is the manufacturer's responsibility to assure that additional production units of this model are manufactured with identical electrical and mechanical components. The manufacturer/importer is responsible to the Competent Authorities in Europe for any modifications made to the production units which result in non-compliance to the relevant regulations. TÜV SÜD Certification and Testing (China) Co., Ltd. Guangzhou Branch shall have no liability for any deductions, inferences or generalizations drawn by the client or others from TÜV SÜD Certification and Testing (China) Co., Ltd. Guangzhou Branch issued reports.

This report is the confidential property of the client. As a mutual protection to our clients, the public and ourselves, extracts from the test report shall not be reproduced except in full without our written approval

TÜV SÜD Certification and Testing (China) Co., Ltd. Guangzhou Branch
5F, Communication Building, 163 Pingyun Rd. Guangzhou 510656 P.R.China, Fax: +86 20 3832 0478 Tel.: +86 20 3832 0668

DIRECTORY - EMISSIONS

			Pages
A)	**Documentation**		
	Directory		2
	Test Regulations		3
	General Remarks and Summary		11
B)	**Test Data**		
	Conducted Emissions	9/150/450 kHz - 30 MHz	5, 10
	Radiated Emissions	30 MHz - 1000 MHz	6, 10
	Interference Power	30 MHz - 300 MHz	7, 10
	Harmonic Current Emissions and Voltage Fluctuations and Flicker	2nd through 40th Harmonics	8, 10
C)	**Appendix A**		
	Test Setup and Test Data Sheets		12 – 17
D)	**Appendix B**		
	Constructional Data Form and Product Information Form(s)		18 – 23
E)	**Appendix C**		
	Constructional Photographs		24 – 30

TÜV SÜD Certification and Testing (China) Co., Ltd. Guangzhou Branch
5F, Communication Building, 163 Pingyun Rd. Guangzhou 510656 P.R.China, Fax: +86 20 3832 0478 Tel.: +86 20 3832 0668

EMISSIONS TEST REGULATIONS :

The emissions tests were performed according to the following regulations:

■ - EMC - Directive 2014/30/EU and its amendments

■ - EN 61326-1:2013

■ - CISPR 11:2009+A1:2010 ■ -Class A, Group 1

Note: For undated references, the latest edition of the publication at the time of testing (including amendments) was applied.

TÜV SÜD Certification and Testing (China) Co., Ltd. Guangzhou Branch
5F, Communication Building, 163 Pingyun Rd. Guangzhou 510656 P.R.China, Fax: +86 20 3832 0478 Tel.: +86 20 3832 0668