Ref. Certif. No.

BE-37040

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Air Purification Sterilizer
Name and address of the applicant	Qingdao Haier Biomedical Co.,Ltd. Haier Industrial Park, Economic Technology Development Zone, Qingdao, China
Name and address of the manufacturer	Qingdao Haier Biomedical Co.,Ltd. Haier Industrial Park, Economic Technology Development Zone, Qingdao, China
Name and address of the factory	Qingdao Haier Biomedical Co.,Ltd. No.280 Fengyuan Road, High-tech Zone, Qingdao (266111), China
Note: *When more than one factory, please report on page 2*	☐ Additional Information on page 2
Ratings and principal characteristics	220 V - 240 V; 50/60 Hz; 100 W; Class II; IPX0
Trademark (if any)	Haier
Customer's Testing Facility (CTF) Stage used	-
Model / Type Ref.	YKJX-Y500
Additional information (if necessary may also be reported on page 2)	- ☐ Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60335-1:2010, IEC 60335-1:2010/AMD1:2013, IEC 60335-1:2010/AMD2:2016 National Differences: EU Group Differences
As shown in the Test Report Ref. No. which forms part of this Certificate	SHES200400698601

This CB Test Certificate is issued by the National Certification Body

SGS Belgium NV - Division SGS CEBEC
Riverside Business Park
Avenue Internationalelaan 55, Building D
B-1070 Brussels, Belgium

Date: 2020-06-19

Signature: Andrew Zhai

This certificate is issued by the company under its General Conditions for Certification Services accessible at http://www.sgs.com/en/Terms-and-Conditions.aspx
Attention is drawn to the limitations of liability defined therein and in the Test Report here above mentioned which findings are reflected in this certificate. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.